

Manufacturer's declaration of conformity

Name and address	Care of Sweden AB Visit: Fabriksgatan 5A, SE-514 33 Tranemo, Mail: Box 146, SE-514 23 Tranemo Phone: +46 (0)771 106 600, Fax +46 (0)325 128 40 info@careofsweden.se
SRN	Not available yet.
Product group	Automatic active airflow controlled mattress systems
Basic UDI-DI	7331345A0010B0010FC
EMDN code	V080701 - ACTIVE ANTI-DECUBITUS MEDICAL MATTRESSES
Classification	Class I, according to Annex VIII (MDR (EU) 2017/745, rule 13
Product/Device name	Control unit CuroCell® IQ
Article/Part number	List of Sales configuration Control unit CuroCell® IQ (IQ-CE-010)
Common Specifications	There are no applicable Common Specifications.


Mark of compliance**Declaration**

We declare under our sole responsibility as Manufacturer that the product(s) listed above conform to the requirements of the MDR (EU) 2017/745. The product(s) meet(s) the relevant General Safety and Performance Requirements of Annex I.

The conformity assessment procedure was performed following Annex II to III of MDR (EU) 2017/745. Any modification to the device, not authorized by us, will invalidate this declaration.

Valid from: 2020-07-01

Updated: 2021-09-29

Manufacturer's signature

Magnus Högberg, CEO

Date 2021-09-29

